

Accuracy of Pulse Oximeters and the Influence of Sensor Type in Induced Hypoxia.

Rochford P.D., Brazzale D.J. Zubrinich C., Pierce R.J. *Proceeding of the America Thoracic Society Annual Meeting*, 2007. A606.

Modern pulse oximeters use advanced signal processing technology to measure arterial oxygen saturation (SpO₂) and may offer improved performance. We evaluated the accuracy of three newer pulse oximeters (Compumedics Vampire (V), Nellcor N595 (N) and Masimo Radical (M) when used with digital sensors in induced hypoxia. A further two N595s were used with ear and forehead reflectance sensors to enable the effect of sensor type to be assessed.

Methods

Simultaneous measurements of SaO₂ by CO-Oximetry and SpO₂ from the five pulse oximeters were obtained in eleven normal subjects over a range of inspired O₂ fractions (FiO₂ =0.10 to 0.21). Hyperoxia was also studied. Bias (mean difference) precision (SD of difference) and root mean square error (RMS) were calculated. The US FDA recommends RMS <3% over SaO₂ range of 70-100%.

Results

212 samples were analyzed with SAO₂ ranging from 71.7 to 100%. Results were similar for all three oximeter types when used with finger probes with RMS <3%. Precision and RMS were markedly better for the reflectance sensor compared with ear or finger sensors. V-finger and N-ear bias was correlated with SaO₂ (p<0.001) and arithmetic correction reduced RMS to 2.3 and 2.2%, suggesting potential for increased accuracy.

Conclusion

Average bias was negligible and RMS was within FDA specification for all pulse oximeter sensor combinations. The forehead reflectance sensor showed superior precision compared with finger and ear sensors.